



**Ground Zero Electrostatics  
MTR-4000 CERTIFICATION AND VERIFICATION**

**CERTIFICATION AND VERIFICATION**

**Model Number: MTR-4000**

**Description:** Portable Wrist Strap/Footwear Tester

Serial Number: 35749

Calibration wd Verification Date: 10/28/04

This is to certify that the above instrument has been calibrated in accordance with our specifications using equipment traceable to NIST per ANSI A540.1

**Tolerances:**

---

**For Single and Dual Wrist Strap:**

<b>Range</b>	<b>Low</b>	<b>High</b>
.5 Meg ohms/10 Meg ohms	500K +-11%	10 Meg +-5%

---

**For Heel Grounders and Conductive Footwear:**

<b>Range</b>	<b>Low</b>	<b>High</b>
.9 Meg ohms/35 Meg ohms	900K +-8%	35 Meg +-6%

---

Note:

It is recommended that the unit be verified every twelve months from date of verification, The unit has no external adjustments and utilizes fixed resistance as shown in the above tolerances.

P. McGowan

Calibrate by: \_\_\_\_\_

